EstÃ¡ndares CAL Checkâ„¢ de Plata â€“ HI97737-11

Description

HI97737-11 is a high quality set of calibration standards that allows users to validate and calibrate the silver range found in the compatible HI97737 portable pHotometer that has a 0.000 to 1.000 mg/L (ppm) silver range. The HI97737-11 includes one sealed ?A? cuvette for zeroing the meter and one ?B? cuvette that is used to validate or calibrate. The expected reading from the ?B? standard is 0.500 ?0.010 mg/L silver @ 25C. El conjunto de calibraciÃ³n se produce en nuestras instalaciones de fabricaciÃ³n de Ãºltima generaciÃ³n utilizando productos quÃmicos de grado reactivo en un entorno de temperatura controlada.

Suministrados con Certificado de AnÃ¡lisis

- CÃ³digo de producto
- NÃºmero de lote
- Fecha de caducidad
- Valor estÃ¡ndar a 25C
- Medidor de referencia trazable al NIST

Contenedores de almacenamiento proporcionados

- A prueba de luz
- Protegidos de roturas accidentales

La validaciÃ³n y calibraciÃ³n del FotÃ³metro PortÃ¡til de Plata HI97737 es rÃ¡pida y fÃ¡cil con los EstÃ¡ndares CAL Check HI97737-11.