



Estándares CAL Check™ de Plata – HI97737-11

## Description

HI97737-11 is a high quality set of calibration standards that allows users to validate and calibrate the silver range found in the compatible HI97737 portable pHotometer that has a 0.000 to 1.000 mg/L (ppm) silver range. The HI97737-11 includes one sealed ?A? cuvette for zeroing the meter and one ?B? cuvette that is used to validate or calibrate. The expected reading from the ?B? standard is 0.500  $\pm$ 0.010 mg/L silver @ 25C. El conjunto de calibración se produce en nuestras instalaciones de fabricación de última generación utilizando productos químicos de grado reactivo en un entorno de temperatura controlada.

## Suministrados con Certificado de Análisis

- Código de producto
- Número de lote
- Fecha de caducidad
- Valor estándar a 25C
- Medidor de referencia trazable al NIST

## Contenedores de almacenamiento proporcionados

- A prueba de luz
- Protegidos de roturas accidentales

La validación y calibración del Fotómetro Portátil de Plata HI97737 es rápida y fácil con los Estándares CAL Check HI97737-11.